



DEPARTMENT OF HEALTH & HUMAN SERVICES

Public Health Service

Food and Drug Administration  
9200 Corporate Boulevard  
Rockville MD 20850

DEC 16 1999

Ms. Danielle Cottet  
Quality Controller  
Wear Safe Malaysia Sdn. Bhd.  
Lot 1, Leboh Hishamuddin Satu  
North Klang Straits Industrial Area  
42000 Port Klang  
Selangor Darul Ehsan, West Malaysia

Re: K992681  
Trade Name: Patient Latex Examination Glove Powdered &  
Powder-Free, Sterile  
Regulatory Class: I  
Product Code: LYY  
Dated: November 25, 1999  
Received: December 2, 1999

Dear Ms. Cottet:

We have reviewed your Section 510(k) notification of intent to market the device referenced above and we have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.


If your device is classified (see above) into either class II (Special Controls) or class III (Premarket Approval), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 895. A substantially equivalent determination assumes compliance with the Good Manufacturing Practice for Medical Devices: General (GMP) regulation (21 CFR Part 820) and that, through periodic GMP inspections, the Food and Drug Administration (FDA) will verify such assumptions. Failure to comply with the GMP regulation may result in regulatory action. In addition, FDA may publish further announcements concerning your device in

the Federal Register. Please note: this response to your premarket notification submission does not affect any obligation you might have under sections 531 through 542 of the Act for devices under the Electronic Product Radiation Control provisions, or other Federal laws or regulations.

This letter will allow you to begin marketing your device as described in your 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801 and additionally 809.10 for in vitro diagnostic devices), please contact the Office of Compliance at (301) 594-4692. Additionally, for questions on the promotion and advertising of your device, please contact the Office of Compliance at (301) 594-4639. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR 807.97). Other general information on your responsibilities under the Act may be obtained from the Division of Small Manufacturers Assistance at its toll-free number (800) 638-2041 or (301) 443-6597 or at its internet address "<http://www.fda.gov/cdrh/dsmamain.html>".

Sincerely yours,

  
for

Timothy A. Ulatowski  
Director  
Division of Dental, Infection Control  
and General Hospital Devices  
Office of Device Evaluation  
Center for Devices and  
Radiological Health

Enclosure

510 (k) Number (if known) : K992681

Device Name : PATIENT EXAMINATION "POWDER" GLOVES STERILE

Indications For Use : Latex

\* A patient Examination Powder glove sterile is a disposable device intended for medical purposes that is worn on the examiner's hand or finger to prevent contamination between patient and examiner.

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

\_\_\_\_\_  
Concurrence of CDRH, Office of Device Evaluation (ODE)

Prescription Use \_\_\_\_\_

OR

Over-The -Counter Use X

(Per 21 CFR 801.109)

(Optional format 1-2-96)

Quinn Lin

(Division Sign-Off)  
Division of Dental, Infection Control,  
and General Hospital Devices

510(k) Number K 992681

Page 01 of 01

510 (k) Number (if known) : K992681

Device Name : PATIENT EXAMINATION "POWDER-FREE" GLOVES STERILE

Indications For Use :

\* A patient Examination Powder-Free Chlorinated glove sterile is a disposable device intended for medical purposes that is worn on the examiner's hand or finger to prevent contamination between patient and examiner.

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

\_\_\_\_\_  
Concurrence of CDRH, Office of Device Evaluation (ODE)

Prescription Use \_\_\_\_\_

OR

Over-The -Counter Use X

(Per 21 CFR 801.109)

(Optional format 1-2-96)

Chin S. Lin

(Division Sign-Off)

Division of Dental, Infection Control,  
and General Hospital Devices

510(k) Number K992681